

Declaration of Conformity

Manufacturer	Atlas Copco Ltd. (trading as Atlas Copco Medical) Unit 18, Nuffield Way, Abingdon, OX14 1RL, UK
Product	Medical Gas Area & Central Alarms
Model Range	PMGS Zeus 6 Area Alarm and Zeus 15 Central Alarm
Classification	IIb
Conformity Route	Annex II
Quality Management System	EN ISO 13485:2012
Quality Management Representative	Ken Tombs – QSHE Manager
GMDN Code	35222, 36271
GMDN Term	Medical Gas and Vacuum Supply Systems
Standards Applied	<i>EN 1041, EN ISO 7396-1, EN ISO 10993-1, EN ISO 14971, EN ISO 15001, EN ISO 15223-1, EN 50581, EN 60601-1, EN 60601-1-2</i>
Notified Body	Lloyd's Register Quality Assurance Limited, 71 Fenchurch Street, London EC3M 4BS United Kingdom (LRQA Notified Body Number 0088)
MDD Certificate(s)	LRQ 4007749/C
Start of CE Marking	03 APR 2013
Place and Date of Issue	Abingdon, 17 May 2018

We hereby declare that the above mentioned products meet the provisions of the Council Directive 93/42/EEC concerning Medical Devices, as amended by Directive 2007/47/EC. All supporting documentation is retained under the premises of the manufacturer.

Endorsing Signature
Sanjay Safaya (General Manager)

