

# Declaration of Conformity

<b>Manufacturer</b>	<b>Atlas Copco Ltd.</b> (trading as Atlas Copco Medical) Unit 18 , Nuffield Way, Abingdon, OX14 1RL, UK
<b>Product</b>	<b>Area Valve Service Unit Modules</b>
<b>Model Range</b>	<b>PMGS Zeus AVSU Modules</b>
<b>Classification</b>	Ila (without alarm) Iib (with alarm)
<b>Conformity Route</b>	Annex II
<b>Quality Management System</b>	EN ISO 13485:2012
<b>Quality Management Representative</b>	Ken Tombs – QSHE Manager
<b>GMDN Code</b>	<b>36271</b>
<b>GMDN Term</b>	<b>Medical Gas and Vacuum Pipeline Systems</b>
<b>Standards Applied</b>	<i>EN 1041, EN ISO 7396-1, EN ISO 10993-1, EN ISO 13348, EN ISO 14971, EN ISO 15001, EN 15223-1, EN ISO 18082, EN 60601-1, EN 60601-1-2, HTM 2022, HTM 02-01</i>
<b>Notified Body</b>	Lloyd’s Register Quality Assurance Limited, 71 Fenchurch Street, London EC3M 4BS United Kingdom (LRQA Notified Body Number 0088)
<b>MDD Certificate(s)</b>	LRQ 4007749/C
<b>Start of CE Marking</b>	<b>03 APR 2013</b>
<b>Place and Date of Issue</b>	<b>Abingdon, 17 May 2018</b>

We hereby declare that the above mentioned products meet the provisions of the Council Directive 93/42/EEC concerning Medical Devices, as amended by Directive 2007/47/EC. All supporting documentation is retained under the premises of the manufacturer.

Endorsing Signature  
Sanjay Safaya (General Manager)

